



## Bassi Logic™ M Manual Files

Heat Treated Manual Nickel Titanium Files

Sizes: 25/.01, 35, 40, 15/.05, 35, 40, 25/.06

Lengths: 21mm & 25mm

Available in 3 procedure kit variations.

### RX Only. For Dental Use Only.

#### Composition

The functional portion (blade) of the instrument is composed of nickel titanium alloy.

#### Indications for Use

Shaping of the root canal system in endodontic procedures.

#### Warnings

Instruments must be sterilized before use. Instruments contain nickel and are contraindicated in individuals with nickel allergies.

#### Contraindications & Adverse Reactions

None Known.

#### Precautions

1. Before clinical use:
  - Consult the IFU.
  - Practice with the files in extracted teeth and plastic blocks.
2. Straight-line access is recommended.
3. Irrigate, recapitulate, and wipe the flutes of the Bassi Logic™ M instruments after every insertion.
4. Bassi Logic™ M must be sterilized prior to clinical use according to ANSI/ADA Specification 28.
5. Dispose of used Bassi Logic™ M instruments in an appropriate Biohazard Sharps container.

### Bassi Logic™ M - Posterior Teeth Clinical IFU

1. Achieve apical patency with a #10 K file.
2. Insert the 25/.06 (Red) file into the coronal and middle canal third until dentin is engaged. Once engaged, begin rotating the file clockwise until the file spins freely.
3. Establish the true working length (TWL)
4. Insert the 25/.01 (White) file until apical patency is obtained using the clockwise rotary motion described in step 2. If patency cannot be obtained with the 25/.01 file, re-establish patency with a #10 and/or possibly #15 K file. Attempt again to achieve patency with the 25/.01 file. Once patency is achieved, move to step 5.
5. Insert the 15/.05 (Yellow) file until dentin is engaged and rotate the file clockwise until it spins freely at TWL.
6. Insert the 25/.06 (Red) file until dentin is engaged and rotate the file clockwise until it spins freely at TWL.
7. After canal preparation, insert the Bassi Clean™ irrigation file to TWL and rotate (or reciprocate) in sodium hypochlorite and again in liquid EDTA for a minimum of 30 seconds each, optimally more.

#### Please note:

- Irrigate and recapitulate after every Bassi Logic™ M file.

### Bassi Logic™ M - Anterior Teeth Clinical IFU

1. Achieve apical patency with a #10 K file.
2. Establish the true working length (TWL)
3. Insert the 35/.01 (White) or 40/.01 (Black) file into the coronal and middle canal third until dentin is engaged and begin rotating the file clockwise until apical patency is obtained. If patency cannot be obtained, re-establish patency with a #10 and/or possibly #15 K file. Attempt again to achieve patency with the 35/.01 or 40/.01 file. Once patency is achieved, move to the next step in the sequence.5.
- 4, Insert the 35/.05 (Green) (or 40/.05 (Black) file until dentin is engaged and rotate the file clockwise until it spins freely at the TWL.
5. After canal preparation, insert the Bassi Clean™ irrigation file to TWL and rotate (or reciprocate) in sodium hypochlorite and again in liquid EDTA for a minimum of 30 seconds each, optimally more.

#### Please note:

- Irrigate and recapitulate after every Bassi Logic™ M file.

